

EC – Declaration of Conformity

| | |
|--|--|
| Manufacturer: | Beltone A/S Lautrupbjerg 7 DK-2750 Ballerup Denmark |
| Conformity Assessment Procedure: | Annex VII of Medical Device Directive (MDD) 93/42/EEC Annex III of Radio Equipment Directive (RED) 2014/53/EU |
| Identification of the Device: | Category: Accessory Type: Assistive Listening Device Brand: Beltone Model: myPAL Pro SM-2P |
| | Revision: Report No. 296 |
| Classification of the Device (MDD): | Class I, Rules 1 and 12, MDD 93/42/EEC |

Applied standards and normative standards:

MDD: EN 60601-1-2:2007/ AC:2010, EN ISO 10993-5:2009, ISO 10993-10:2010,
EN ISO 14971:2012
RED: Health & Safety: EN 60950-1:2006 + A2:2013, EN/(IEC) 62479:2010;
EMC: EN 301 489-17 V3.2.0, EN 301 489-1 V2.2.0;
Spectrum: EN 300 328 V2.1.1, EN 300 422-4 V2.1.1.

We, the manufacturer hereby declare that the above-mentioned devices comply with the relevant provisions of the EU Council Directive 93/42/EEC (MDD) Annex I - Essential Requirements and the essential requirements and other relevant requirements of the RED 2014/53/EU and their relevant transpositions into national laws of the Member States in which the above mentioned medical devices are distributed.

Place and Date: Ballerup, 26 October 2017



Helen Ljungdahl Round
Senior Vice President
Global Marketing
Beltone A/S



Lars Hagander
Vice President
Corporate Quality
Beltone A/S

Doc. ref. 00066490