Beltone Remote Care Live

Quick guide

Before your Beltone Remote Care Live appointment, please ensure:

- Your have downloaded the latest version of the Beltone HearMax™ app and your hearing aids are paired to your iPhone
- You are in a location with a strong Wi-Fi connection.
- You insert fresh batteries into your hearing aids. If you have rechargeable hearing aids, make sure they are fully charged.
- You have at least 50% battery on your iPhone.
- Audio routing is set to 'Automatic' to hear audio through the hearing aids or set to 'Never Hearing Devices' to
 hear audio through the phone speaker. To check your settings open Settings >Accessibility > Hearing Devices
 > Audio Routing > Call Audio > select Automatic OR Never Hearing Devices.

For a full readiness checklist, open the Beltone HearMax app > My Beltone > Live Assistance.

- 1. Your hearing care professional will initiate the Beltone Remote Care Live call. If your iPhone is locked, swipe to answer the call or tap **Accept**.
- 2. Tap the Beltone HearMax icon to enable video access.
- 3. Once connected, you will see your hearing care professional on the screen. While your hearing care professional is making adjustments to your hearing aids, you will hear his or her voice through your iPhone speaker instead of your hearing aids. At any time, you can select any of the available icons to initiate a chat, turn on or off audio and video or hang up the call.
- 4. You will see a progress bar appear as your hearing aids are being updated.
- 5. Once the session has ended, your hearing aids will automatically reboot and you will receive a 'Session ended' notification on your iPhone. Tap Close to enjoy your new settings.

Your hearing care professional







M102786GB-20.04-Rev.A