Made for iPhone | iPad | iPod

Works with android

User guide

Beltone Behind-The-Ear hearing aids

Beltone Rely™

Beltone

Hearing aid information

Left hearing aid		Right hearing aid	
Serial number		Serial number	
Model number		Model number	
Battery type	☐ Mini BTE Size 312 ☐ Standard BTE Size 13		

Program	Beep	Description
1	One beep	
2	Two beeps	
3	Three beeps	
4	Four beeps	

Table of Contents

Introduction 5
Your hearing aid 7
How to get your hearing aid ready for use 9
How to place the hearing aids in your ears 13
How to remove the hearing aids from your ears 15
How to use your hearing aids 16
Direct Audio Input 25
Advanced options 27
How to clean and maintain your hearing aids 32
General warnings and precautions 38
Hearing aid expectations 40
Troubleshooting 41
Tinnitus Management 43
Regulatory information 52
Technical specifications 57
Additional information 59

Introduction

Thank you for choosing Beltone hearing aids. We recommend that you use your hearing aids every day. This way you will fully benefit from them.

NOTE: Read this booklet carefully before you start using your hearing aids.

Intended use

Generic air-conduction hearing aids are wearable sound-amplifying devices intended to compensate for impaired hearing. The fundamental operating principle of hearing aids is to receive, amplify, and transfer sound to the eardrum of a hearing impaired person.

Contra-indications

A hearing care professional should advise a prospective hearing aid user to consult promptly with a licensed physician (preferably an ear specialist) before dispensing a hearing aid, if the hearing aid dispenser determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- Visible congenital or traumatic deformity of the ear.
- History of active drainage from the ear within the previous 90 days.
- History of sudden or rapidly progressive hearing loss within the previous 90 days.
- Acute or chronic dizziness.
- Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- Audiometric air-bone gap equal to or greater than 15 dB at 500 Hertz (Hz), 1000 Hz, and 2000 Hz.

- Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- Pain or discomfort in the ear.

Side effects

If you experience side effects, contact your hearing care professional. Possible side effects from wearing a hearing aid may be:

- Dizziness
- Tinnitus
- Perceived worsening of hearing loss
- Nausea
- Skin reaction
- Ear wax accumulation

For devices including a Tinnitus Breaker Pro module
The Tinnitus Breaker Pro module is a tool to generate sounds to be used in a Tinnitus Management Program to temporarily relieve patients suffering from Tinnitus.

For devices including a dome
The dome is intended to be connected to a thin tube on the hearing aid. The dome is intended to ensure that the sound outlet of the hearing aid is placed in the ear canal.
This accessory is intended to be used by the same age group as the hearing aid. The accessory is intended to be used by lay persons.

Your hearing aid

76 model

1. Thin tube
2. Open dome
3. Microphone inlets
4. Push button
5. Volume control
6. Battery compartment
7. Direct Audio Input
8. Battery door lock (optional)



The hearing aids in the illustrations are shown with a thin tube and an open dome, but they can also be fitted with other types of domes/earmoulds:



Use only original consumables from Beltone, e.g. tubes and domes.

66 model

1. Thin tube
2. Open dome
3. Push button
4. Microphone inlets
5. Battery compartment



The hearing aids in the illustrations are shown with a thin tube and an open dome, but they can also be fitted with other types of domes/earmoulds:



Use only original consumables from Beltone, e.g. tubes and domes.

How to get your hearing aid ready for use

Battery warnings

WARNING: Batteries contain dangerous substances and should be disposed of carefully in the interest of your safety and for the environment. Please note:

- Keep hearing aid batteries away from pets, children, and people with cognitive, intellectual or mental health challenges.
- Never put a hearing aid battery in your mouth.
- Never swallow batteries nor place them inside any part of the body, as the battery can cause serious injuries. If a battery has been swallowed or placed inside any part of the body, seek immediate medical attention.
- Do not recharge zinc-air batteries – they may leak or explode.
- Do not attempt to dispose of batteries by burning them.
- Used batteries are harmful to the environment. Please dispose of them according to local regulations or return them to your hearing care professional.
- Batteries may leak. Remove the battery if you leave the hearing aids unused for longer periods.
- If the batteries are not inserted correctly, the device will not work and the batteries may build up heat. If this happens, please remove the batteries.

NOTE:

- Always use new zinc-air batteries that have a minimum remaining shelf life of one year.
- To save battery power, turn off your hearing aids when they are not in use.

How to change the battery

1. Prepare the new battery. Remove the protective foil to activate the battery - wait for **two minutes** before inserting the battery into the hearing aid.

2. Open the battery door completely using your fingernail.
3. Remove the used battery.



4. Insert the new battery with the positive side (+) facing upwards. Always insert the battery in the door, never directly into the hearing aid.
5. Close the battery door.



NOTE: Please observe the following:

- To save battery power, turn off your hearing aids when they are not in use.
- At night, switch off the hearing aid and open the battery door completely to allow moisture to evaporate. This prolongs the hearing aid's lifespan.
- If the hearing aid frequently loses connection with wireless accessories, contact your hearing care professional for a list of appropriate batteries.

Battery door lock

If the hearing aid is going to be used by a child or a mentally disabled person, you can ask your hearing care professional to make a battery door lock available for you. You can turn the hearing aid on and off as you usually do, but you will have to unlock the battery door every time you need to change the battery.

How to use the battery door lock

NOTE: The battery lock is not available for 66 models.

Low battery warning

When the batteries are low on power, your hearing aids reduce the volume, and play a melody every 15 minutes, until they are empty and turn off.

NOTE: Keep spare batteries on hand.

Low battery alert when paired with wireless accessories (optional)

NOTE:

- Your batteries will drain faster when you use wireless functions like streaming from your smart device or from your TV with our TV streamer. As the battery power declines, the wireless functions stop working. A short melody will play every five minutes to let you know that the battery power is low. The table below shows the functionality with different battery charge levels.
- If the hearing aids are experiencing frequent loss of connection to wireless accessories, contact your hearing care professional for a list of low impedance batteries.

Battery level	Signal	Hearing aid	Remote control	Streaming
Fully charged		✓	✓	✓
Low	4 even tones	✓	✓	x
Depleted	3 even tones and 1 longer tone	✓	x	x

These will work again when you insert a new battery.

How to place the hearing aids in your ears

How to tell left from right

If you have two hearing aids, they may be programmed differently. One for your left ear, the other for your right. Do not swap them. Please pay attention to this when cleaning, storing, and inserting the hearing aids.

You might want to ask your hearing care professional to mark your hearing aids with a coloured Left and Right indication: Left is blue and Right is red.

Inserting earmoulds into your ears

1. Hold the earmould between your thumb and index finger and position its sound outlet in your ear canal.
2. Slide the earmould all the way into your ear with a gentle, twisting movement.
3. Turn the top part of the earmould gently backwards and forwards so it tucks behind the fold of skin above your ear canal. Move the earmould up and down and gently press it to place it correctly in the ear.
4. Place the hearing aid firmly behind the ear and make sure it sits securely. When properly inserted, your hearing aids should fit snugly and comfortably.

NOTE: It may be helpful to pull your ear up and outward with your opposite hand during insertion. By experimenting, you may discover an easier method.

CAUTION: Never attempt to modify the shape of the hearing aid, earmoulds or tubing yourself.

How to remove the hearing aids from your ears

1. Lift the hearing aid from behind the ear. Let it hang momentarily beside your ear
2. Using your thumb and index finger, take hold of the earmould (not the hearing aid or the tubing)
3. Gently twist and pull the earmould to remove it from the ear

How to use your hearing aids

Turning your hearing aids on and off

Once you have placed the hearing aids on your ears, you can turn them on.

Your hearing aids always start in program 1 at the pre-set volume.



To turn your hearing aid on, close the battery door.



To turn the hearing aid off, open the battery door (with your fingernail).

How to adjust the volume

Your hearing aid automatically adjusts the volume depending on your listening situation.

However, if your hearing aid has a volume control, you can adjust the volume according to your preferences.

1. To increase the volume, briefly press the **top** part of the button.
2. To reduce the volume, briefly press the **bottom** part of the button.



When you change the volume, the hearing aid responds with a beep. When you reach the upper or lower limits, the hearing aid responds with a low-pitched beep.

NOTE: Not available for 66 models.

You can also adjust the volume by means of your Beltone Direct Remote Control 2 or the Beltone HearMax™ app.

NOTE:

- If you have two hearing aids with the Synchronised Volume Control function enabled, volume control adjustments to one hearing aid automatically repeat in the second hearing aid. When you change the volume in one of the hearing aids, it responds with one or more beeps. A beep in the second hearing aid follows.
- Your hearing care professional can disable the volume control or replace it with a non-functional cover.

How to change program

Your hearing aid has a push button which allows you to select from several listening programs.

Push the button to change program. You will then hear one or more beeps. The number of beeps indicates which program you have selected (one beep = program 1, two beeps = program 2 and so on).



You can also change programs using a Beltone Direct Remote Control 2 or the Beltone HearMax™ app.

NOTE:

- If you have two hearing aids with synchronisation enabled, changing the program on one hearing aid automatically applies the change to the second hearing aid as well. The same number of beeps will then sound in the second hearing aid.

- When you turn the hearing aids off and on again, they always start up in program 1 and your pre-set volume level.

Telecoil

Not available for model RLY176-DW.

Your hearing aid may have a telecoil. The Telecoil function may help to improve understanding of speech with Hearing Aid Compatible (HAC) telephones and in theatres, cinemas, houses of worship, etc. that have a teleloop installed.

When you select the Telecoil program, your hearing aid picks up signals from the teleloop or HAC telephone. Your hearing care professional can activate the Telecoil program.

NOTE:

- The telecoil does not work without a teleloop (that is, an induction loop) or a HAC telephone.
- If you are having trouble hearing with the telecoil, ask you hearing care professional to adjust the program.
- If there is no sound from your hearing aids in a teleloop system with an active Telecoil function, the teleloop system may not be turned on or may not be operating correctly.
- The sound from the teleloop and the hearing aids' microphones can be mixed according to your preference. Ask your hearing care professional for more details.

Teleloop systems

To use teleloop systems, follow these steps:

1. Switch your hearing aid to the Telecoil program.
2. Find a good spot. Reception is not clear in all locations, it depends on the teleloop. Look for signs or find another spot .
3. If needed, adjust the volume.
4. When you leave, switch to program 1.

HAC telephone

Some smartphones are hearing aid compatible (HAC). The HAC phone establishes a small hearing loop that your hearing aids can connect to. The telecoil picks up the HAC phone's signal and converts it to sound.

To use a HAC phone, follow these steps:

1. Switch your hearing aid to the Telecoil program.
2. Pick up the phone and make a call or answer a call.
3. Hold the phone close to the hearing aid and tilt it slightly outwards.
4. Listen to the dial tone and move the telephone to get the best reception.
5. If needed, adjust the volume.
6. When you hang up, switch back to your preferred program.

NOTE:

- If the phone has a poor telecoil signal, use the microphone program. To avoid whistling, do not hold the handset too tightly against your ear.
- Ask your hearing care professional to enable the Telecoil programme in your hearing aids.
- If you see a “M3”, “M4”, “T3”, or “T4” on the box, then the smartphone is HAC compliant. If you find it difficult to obtain a good result while using your smartphone, your hearing care professional will be able to give you advice on available wireless accessories to enhance listening capabilities. Ask your hearing care professional for advice regarding HAC smartphones.

Using a telephone



Your hearing aid allows you to use your telephone as you normally do. Finding the optimal position for holding the phone may require practice.

The following suggestions may be helpful:

1. Hold the telephone up to your ear canal or hold it close to the hearing aid microphones as illustrated.
2. If you hear whistling, try holding the telephone in the same position for a few seconds. The hearing aid may be able to cancel the whistling.
3. You can also try holding the telephone slightly away from the ear.

NOTE: Depending on your needs, your hearing care professional may activate a function specifically for telephone use.

Mobile phones

Your hearing aids comply with the most stringent Standards of International Electromagnetic Compatibility. Any degree of disturbance can be due to the nature of your particular mobile phone or of your wireless telephone service provider.

NOTE:

- If you find it difficult to get a good result while using your mobile phone, your hearing care professional can give you advice on available wireless accessories to enhance listening capabilities. See also "Wireless accessories", page 30.

Auto-Phone (optional)

If you place a magnet on the telephone receiver, your hearing aids will automatically switch the telephone program on when the receiver is close to your ear. When you remove the receiver from your ear, the hearing aids automatically return to the previous listening program.

NOTE: Ask your hearing care professional to enable Auto-Phone as one of your programs.

⚠ Auto-Phone warnings

- If a magnet is swallowed, seek immediate advice from a medical practitioner.
- Keep magnets out of reach of pets, children and mentally disabled persons.
- The Auto-Phone magnet may affect sensitive medical devices/electronic systems. Seek advice from the manufacturers regarding appropriate safety measures when using the Auto-Phone solution near the sensitive

device/equipment (pacemakers and defibrillators) in question. If the manufacturer cannot issue a statement, we recommend keeping the magnet or a telephone equipped with the magnet 30 cm (12”) away from magnetically sensitive devices (e.g. pacemakers).

⚠ Auto-Phone precautions

- If you experience frequent signal loss or noise during calls, move the magnet to another place on the telephone receiver.
- Only use magnets supplied by Beltone.

Placing the Auto-Phone magnet

Place the magnet on your telephone receiver as follows:



1. Clean the surface thoroughly. Use a recommended cleaning agent.



2. Remove the foil from the magnet.



3. Place the magnet on the phone.

CAUTION:

- If you experience frequent signal loss or noise during calls, move the Auto-Phone magnet to another place on the telephone receiver.
- Only use magnets supplied by Beltone.

How to use Auto-Phone

1. Lift the telephone to your ear.
2. When you hear a short melody, the phone program is active.

NOTE:

- You may need to move the telephone receiver slightly to find the best position for a reliable Auto-Phone activation and a good hearing experience on the telephone.
- If your hearing aids have enabled the Asymmetric Phone Handling functionality, the hearing aid on the non-phone ear automatically attenuates.
- Do not cover the phone loudspeaker opening with the magnet.
- If the function does not work to your satisfaction, moving the magnet to another position may improve ease of use and comfort.
- If your hearing aids do not switch to the telephone program consistently, try repositioning the magnet or adding additional magnets.
- Use a recommended cleaning agent.

Direct Audio Input

(Optional for model 76)

Not available for model RLY176-DW.

You can connect a DAI (Direct Audio Input) adapter to the bottom of your hearing aid. Once connected, the hearing aid automatically switches to DAI. The sound is then sent directly to your hearing aid using a cable or a wireless FM system.

If you want to be able to hear what happens around you, you can combine the DAI input with the sounds picked up by your hearing aid's microphones.

NOTE: Your hearing aid's battery will drain faster if you use the DAI functionality.

Connecting a DAI adapter

1. Align the tip of your DAI adapter with the groove on top of the battery door.

2. Move the adapter towards the battery door.



3. Click the adapter onto the hearing aid.

Disconnecting a DAI adapter



- Press the small latch downwards and gently remove the adapter from the hearing aid.

⚠ Important points for FM

- Do not use two transmitters on the same FM channel.
- Do not use water or fluids for cleaning the FM click-on (DAI) receiver.
- Do not use an FM transmitter in locations where it is forbidden to use electronic devices, for instance in airplanes/oil rigs.
- Be aware that FM signals might also be picked up and overheard by other receivers.
- Before using the system in another country, contact your hearing care professional to make sure your radio channel is permitted in that country.
- Your FM boot and transmitter may only be repaired by an authorised service centre.

Advanced options

Beltone Remote Care and Beltone Remote Care Live (optional)

Beltone Remote Care
If you have signed up to use Beltone Remote Care available with your hearing aids, you can allow your hearing aids to be adjusted remotely without having to visit your hearing care professional.

All you need is a compatible mobile device with internet enabled. This allows you to:

- Request assistance remotely to adjust your hearing aids to be a better fit for you.
- Keep your hearing aids up to date with the latest software to ensure the best performance possible.

This service only works if your mobile device is connected to the internet. Your hearing care professional will provide information regarding this option, and how it works with the Beltone HearMax™ app.

For optimum performance, make sure the hearing aids are connected to the Beltone HearMax™ app and placed close to the iPhone, iPad, iPod touch, or the Android™ smartphone before applying the changes.

NOTE: Your hearing aids shut down during the installation and update process.

Beltone Remote Care Live
This service also includes Beltone Remote Care Live. With this service you can get face-to-face assistance from your hearing care professional from home.

Using your hearing aids with iPhone, iPad, and iPod touch (optional)

The advanced models of our hearing aids are Made for iPhone, iPad and iPod touch, which allow for direct audio streaming and control from these devices.

Streaming from an Android™ smartphone

Some Android smartphones can stream audio directly to the advanced models of our hearing aids. Your device must be running Android 10 or newer and it must have the Android Streaming for Hearing Aids feature as well.

> **NOTE:** For assistance with pairing and using these products with your hearing aids, contact your hearing care professional.

Using your hearing aid with smartphone apps (optional)

Our smartphone apps are intended to be used with our wireless hearing aids. The smartphone apps send and receive signals from the hearing aids via smartphones.

- Do not disable app notifications
- Install updates to keep the app working correctly
- Only use the app with hearing aids from the same manufacturer. We take no responsibility if the app is used with other hearing aids
- If you want a printed version of the smartphone app user guide, please go to our website at beltone.com or consult customer support

NOTE: For assistance with pairing and using these products with your hearing aids, please contact your hearing care professional or visit our support site.

NOTE: If your Bluetooth® enabled Android smartphone does not stream directly to your hearing aids, you are able to answer the telephone if you use Beltone Direct Phone Link 2.

Download on the App Store

GET IT ON Google Play

Flight Mode (optional)

Your hearing aids can be controlled from your mobile device or Remote Control, an option that can be added by your hearing care professional. However, in some areas you are requested to turn off wireless communication.

CAUTION: When boarding a flight or entering an area where RF transmitters are prohibited, wireless functionality must be deactivated.

NOTE: Both hearing aids must be set in Flight mode - even with synchronisation enabled.

Turning off wireless communication (enter Flight Mode)

1. Open and close the battery door on each hearing aid three times within 10 seconds.
2. A 10-second double tone (♫♫) means the hearing aid is now in Flight Mode.

NOTE: Both hearing aids must be set in Flight mode - even with synchronisation enabled.

Activating wireless communication (exit Flight Mode)

1. Open and close the battery door on each hearing aid once.
2. Wireless communication will be activated after 10 seconds.

NOTE: It is important to wait an additional 15 seconds after wireless function resumes before opening and closing the battery compartment again for any reason. Flight mode will resume if you open and close the battery compartment during this 15 second window.

Wireless accessories

Our wireless eco-system features a comprehensive range of seamlessly integrated wireless accessories. These let you stream high quality stereo sound and speech directly to your hearing aids.

- **A TV streamer** streams audio from a TV and most other audio sources to your hearing aids at a volume that suits you.
- **A basic remote control** adjusts volume, mutes your hearing aids and changes programs.
- **An advanced remote control** adjusts volume, mutes your hearing aids, changes programs and displays your settings.

- **A phone clip** streams phone conversations and stereo sound to both hearing aids and doubles as a remote control.
- **A body-worn microphone** is a microphone that can be worn by others. It improves speech comprehension in noisy situations.
- **A wireless microphone.** It works like the **body-worn microphone**, but doubles as a table microphone. Furthermore, it has a built-in telecoil that allows it to connect with a teleloop system, a connector for a FM receiver, and a mini-jack input for wired streaming of audio from a computer or music player.

Accessing wireless accessories

To access a wireless accessory that has already been paired with your hearing aids, press the push button for 2 seconds. The hearing aid will emit a sound to confirm the connection.

For information about how to pair your hearing aids with a wireless accessory, see the user guide for the relevant wireless accessory.

NOTE:

- Ask your hearing care professional for more information on the range of Beltone wireless accessories.
- You should only use Beltone wireless accessories with your wireless hearing aids.

How to clean and maintain your hearing aids

Care and maintenance

Please follow the advice below to have the best user experience and to prolong the life of your hearing aids.

- Keep your hearing aids dry and clean.
- Open the battery door to dry your hearing aids when you are not wearing them.
- Wipe the hearing aids with a soft cloth after use to remove grease or moisture.
- Do not wear your hearing aids when putting on cosmetics, perfume, aftershave, hairspray, suntan lotion, etc. These might discolour the hearing aid or get into the hearing aid causing damage.
- Do not immerse your hearing aid in any liquid.
- Keep your hearing aids away from excessive heat and intense direct sunlight. The heat may deform the shell, damage the electronics and deteriorate the surface.
- Do not swim, shower or steam bathe while wearing your hearing aids.

Daily maintenance

It is important to keep your hearing aid clean and dry. On a daily basis, clean the hearing aids using a soft cloth or tissue. In order to avoid damage due to humidity or excessive perspiration, the use of a drying kit is recommended.

WARNING: Always turn your hearing aids off while cleaning and maintaining them.

Cleaning tools

1. Soft cloth.
2. Brush for cleaning. Use the brush on all surfaces and orifices. Also use your brush for daily cleaning and battery handling.
3. Wire loop. Use the wire loop to clean the earmould.
4. Magnet. Use the magnet to lift and replace the battery.



If the microphone inlets are clogged, gently brush across the microphone inlets with the brush.



WARNING: Do not use force to press the bristles on the small brush into the inlets because the microphones may be damaged.

CAUTION: Do not use alcohol or other solvents to clean your hearing aid, the protective coating will be damaged.

NOTE: Do not use the wire loop to clean the microphone inlets. If the microphone inlets remain clogged after brushing the exterior, ask a hearing care professional to help you clean them.

NOTE: The wire loop is only intended for earmoulds.

The earmould

1. Remove the earmould and tubing from the hearing aids prior to cleaning.
2. Clean the earmould using a mild soap, and rinse with lukewarm water.
3. After cleaning, dry the earmoulds thoroughly and remove any residual water and debris from the tubing utilising a small blower and wire loop.





NOTE: Your earmould tubing may become stiff, brittle or discoloured over time. Contact your hearing care professional regarding tube changes.

Thin tubes and domes

1. Remove the thin tubes from the hearing aids by unscrewing them counter-clockwise.
2. Wipe down thin tubes and domes with a damp cloth.
3. In order to clear the thin tube of moisture and debris, push the black cleaning wire through the thin tube, beginning at the opposite end of the dome.



NOTE: We recommend that you change the thin tube and dome systems every three months. If the components get stiff or brittle, change them sooner.

How to change domes

1. Remove the used dome by pulling it off the ribbed flange and discard. This may require a bit of force.
2. Push the new dome over the ribbed flange on the thin tube
3. Make sure that the new dome is properly and securely mounted
4. Check that the dome is securely mounted: Carefully lift the lower part of the dome and verify that the collar completely covers the ribbed flange on the tube.





NOTE: This procedure shows an open dome but it is also applicable if your hearing aid has another type of dome.

CAUTION: Use only original consumables from Beltone, e.g. tubes and domes.

General warnings and precautions

⚠ General warnings

1. Consult a hearing care professional if you think there may be a foreign object in your ear canal, if you experience skin irritation, or if excessive earwax accumulates with the use of the hearing aid.
2. Different types of radiation, from e.g. NMR, MRI, or CT scanners, may damage hearing aids. It is recommended not to wear hearing aids during these or other similar procedures. Other types of radiation, such as burglar alarms, room surveillance systems, radio equipment, mobile telephones contain less energy and will not damage hearing aids. However, they have the potential to momentarily affect the sound quality or temporarily create undesired sounds from the hearing aids.
3. Do not wear hearing aids in mines, oil fields, or other explosive areas unless those areas are certified for hearing aid use.
4. Do not allow others to use your hearing aids.
5. Hearing aid usage by children or mentally disabled persons should be supervised at all times to ensure their safety. The hearing aid contains small parts that could be swallowed by children. Please be mindful not to leave children unsupervised with this hearing aid.
6. Hearing aids should be used only as prescribed by your hearing care professional. Incorrect use may result in sudden and permanent hearing loss.
7. Warning to hearing care professionals: Special care should be exercised in selecting and fitting hearing aids with maximum sound pressure level that exceeds 132 dB SPL with an IEC 60711:1981 occluded ear simulator. There may be a risk of impairment of the remaining hearing.

8. Turn off your wireless functionality by using the flight mode in areas where radio frequency emission is prohibited.
9. If a hearing aid is broken, do not use it.
10. A power hearing aid can produce very loud sound to compensate for a severe or profound hearing loss. There is therefore a risk of further impairing the remaining hearing.
11. External devices connected to the electrical input must be safe according to the requirements of IEC 60601-1, IEC 60065, EN/IEC 62368-1, or IEC 60950-1, as appropriate (wired connection, for example HI-PRO, SpeedLink).

NOTE:

For use of wireless functionality, only use supported wireless accessories. For further guidance regarding pairing etc., please refer to the user guide of the relevant wireless accessory.

General precautions

- When wireless function is activated, the device uses low-powered digitally coded transmissions in order to communicate with other wireless devices. Although unlikely, nearby electronic devices may be affected. In that case, move the hearing aid away from the affected electronic device
- Use only original consumables from Beltone, e.g. tubes and domes.
- Only connect your hearing aids to accessories intended and qualified to be used with your hearing aids.

Hearing aid expectations

- A hearing aid will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions.
- Consistent use of the hearing aid is recommended. In most cases, infrequent use does not permit you to get full benefit from it.
- The use of a hearing aid is only part of hearing rehabilitation and may need to be supplemented by auditory training and instructions in lip-reading.

Troubleshooting

Issue	Potential cause	Potential solution
Feedback, "whistling"	Is your earmould or dome inserted correctly in the ear?	Re-insert it.
	Is the volume very loud?	Reduce the volume.
	Are you holding an object (e.g. a hat or a phone) close to a hearing aid?	Move your hand away to create more space between the hearing aid and the object.
	Is your ear full of wax?	Visit your physician.
No sound	Is the hearing aid turned off?	Turn it on.
	Is the hearing aid in telecoil mode?	Switch to the microphone program.
	Is there a battery in the hearing aid?	Insert a new battery
	Is the battery still good?	Replace with a new battery
	Is the plastic tube, earmould or dome clogged or broken?	Consult your hearing care professional.
	Is your ear full of wax?	Visit your physician.

Issue	Potential cause	Potential solution
Sound is distorted, spluttering or weak	The battery is dead	Replace it with a new one
	Is the battery dirty?	Clean it or replace it with a new one
	Is the plastic tube or earmould clogged or broken?	Consult your hearing care professional.
	Did your hearing aid get moist?	Use a desiccant (drying kit).
Battery drains very quickly	Did you leave your hearing aid turned on for long periods of time?	Always turn off your hearing aid when you are not using it, e.g. at night.
	Is the battery old?	Check the date on the battery package.

Tinnitus Management

Tinnitus Breaker Pro module

Not available for model RLY176-DW.

Your Beltone hearing aids include the Tinnitus Breaker Pro (TBP) module. The Tinnitus Breaker Pro (TBP) Module is a software tool that generates sounds to be used in tinnitus management programs to relieve suffering from tinnitus. The TBP can generate sounds adjusted to your personal preference and your specific therapeutic needs as determined by your doctor, audiologist or hearing care professional. Depending on the selected hearing aid program and the environment you are in, you may hear the therapeutic sound of a continuous or fluctuating noise.

Indications for use of the TBP module - (US only)

The Tinnitus Breaker Pro module is a tool to generate sounds to be used in a Tinnitus Management Program to temporarily relieve patients suffering from Tinnitus. The target population is primarily the adult population over 18 years of age. This product may also be used with children 5 years of age or older.

TheTinnitus Breaker Pro Module is targeted for healthcare professionals, which are treating patients suffering from Tinnitus, as well as conventional hearing disorders. The fitting of the Tinnitus Breaker Pro Module must be done by a hearing professional participating in a Tinnitus Management Program.

If deemed feasible by the hearing professional, subsequent fittings of the Tinnitus Breaker Pro module may be performed remotely and in real time while having live communication via live audio, video and chat on the user's dedicated app.

User instructions for the TBP module

Description of the device

The Tinnitus Breaker Pro (TBP) Module is a software tool that generates sounds to be used in tinnitus management programs to temporarily relieve suffering from tinnitus.

Explanation of how the device works

The TBP module is a frequency and amplitude shaped white-noise generator. The noise signal level and frequency characteristics can be adjusted to the specific therapeutic needs as determined by your doctor, audiologist or hearing care professional.

Your doctor, audiologist or hearing care professional can modulate the generated noise with the purpose of making it more pleasant. The noise can then resemble, for example, breaking waves on a shore.

Modulation level and speed can also be configured to your likes and needs. An additional feature can be enabled by your hearing care professional that allows you to select predefined sounds that simulate sounds from nature, such as breaking waves or running water.

If you have two wireless hearing aids that support ear-to-ear synchronisation, this functionality can be enabled by your hearing care professional. This will cause the Tinnitus Breaker Pro to synchronise the sound in both hearing aids.

If your tinnitus troubles you only in quiet environments, your doctor, audiologist or hearing care professional can set the TBP Module so that it becomes audible exclusively in such surroundings. The overall sound level can be adjusted via a volume control. Your doctor, audiologist or hearing care professional will review with you the need for having such a control.

For hearing aids where ear-to-ear synchronisation is enabled, your hearing care professional can also enable environmental monitoring synchronisation so that the TBP noise level is automatically adjusted simultaneously in both hearing aids dependent on the background sound level. Additionally, since the hearing aid has a volume control, the background noise level monitored by the hearing aid and the volume control can be used simultaneously to adjust the generated noise level in both hearing aids.

The scientific concepts that form the basis for the device

The TBP module provides sound enrichment with the aim of surrounding the tinnitus sound with a neutral sound which is easily ignored. Sound enrichment is an important component of most approaches to tinnitus management, such as Tinnitus Retraining Therapy (TRT).

To assist habituation to tinnitus, this needs to be audible. The ideal level of the TBP module, therefore, should be set so that it starts to blend with the tinnitus, and so that you can hear both your tinnitus as well as the sound used.

In a majority of instances, the TBP module can also be set to mask the tinnitus sound, so to provide temporary relief by introducing a more pleasant and controllable sound source.

TBP volume control

The sound generator is set to a specific loudness level by the hearing care professional. When switching the sound generator on, the volume will have this optimal setting. Therefore, it might not be necessary to control the volume (loudness) manually. However, the volume control provides the ability to adjust the volume, or amount of stimulus, to the liking of the user. The tinnitus sound generator volume can only be adjusted within the range set by the hearing care professional.

The volume control is an optional feature in the TBP module used for adjusting the sound generator output level.

Using TBP with smartphone apps

The tinnitus sound generator control via hearing aid push buttons can be enhanced with wireless control from a TBP control app on a smartphone or mobile device. This functionality is available in supported hearing aids when a hearing care professional has enabled the TBP functionality during fitting of the hearing aid.

> **NOTE:** To use smartphone apps, the hearing aid must be connected with the smartphone or mobile device.

TBP - Technical specifications

Audio signal technology

Digital.

Available sounds

White noise signal which can be shaped with the following configurations:

High-pass filter	Low-pass filter
500 Hz	2000 Hz
750 Hz	3000 Hz
1000 Hz	4000 Hz
1500 Hz	5000 Hz
2000 Hz	6000 Hz
-	8000 Hz

The white noise signal can be modulated in amplitude with an attenuation depth of up to 14 dB.

⚠ Prescription use of a Tinnitus Breaker Pro hearing aid

The TBP should be used as prescribed by your doctor, audiologist or hearing healthcare professional. In order to avoid permanent hearing damage, the maximum daily usage depends on the level of the generated sound.

To adjust TBP, please consult your hearing care professional.

Should you develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function or increase in tinnitus perception, you should discontinue use of the sound generator and seek medical evaluation.

Children and physically or mentally challenged users will require training by a doctor, audiologist, hearing care professional or the guardian for the insertion and removal of the hearing aid containing the TBP module.

Important notice for prospective sound generator users

A tinnitus masker is an electronic device intended to generate noise of sufficient intensity and bandwidth to mask internal noises. It is also used as an aid in hearing external noises and speech.

Good health practice requires that a person with a tinnitus condition have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before using a sound generator. Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists or otorhinolaryngologists.

The purpose of medical evaluation is to assure that all medically treatable conditions that may affect tinnitus are identified and treated before the sound generator instrument is used.

The sound generator instrument is a tool to generate sounds to be used with appropriate counselling and/or in a tinnitus management program to relieve patients suffering from tinnitus.

Warning information

WARNING:

- Sound generators can be dangerous if improperly used.
- Sound generators should be used only as advised by your doctor, audiologist, or hearing care professional.

- Sound generators are not toys and should be kept out of reach of anyone who might cause themselves injury (especially children and pets).

CAUTION:

- Should the user develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function or increase in tinnitus perception, the user should discontinue use of the sound generator and seek medical evaluation.
- To prevent unintended usage by paediatric or physically or mentally disabled users, the volume control must, if enabled, be configured to only provide a decrease of the sound generator output level.
- Children and physically or mentally disabled users will require guardian supervision while wearing the TBP hearing aid.

Tinnitus Breaker Pro warning to hearing care professionals

A hearing care professional should advise a prospective sound generator user to consult promptly with a licensed physician (preferably an ear specialist) before getting a sound generator, if the hearing care professional determines through inquiry, actual observation, or review of any other available information concerning the prospective user that the prospective user has any of the following conditions:

1. Visible congenital or traumatic deformity of the ear.
2. History of active drainage from the ear within the previous 90 days.
3. History of sudden or rapidly progressive hearing loss within the previous 90 days.
4. Acute or chronic dizziness.

5. Unilateral hearing loss of sudden or recent onset within the previous 90 days.
6. Audiometric air-bone gap equal to or greater than 15 dB at 500 Hertz (Hz), 1000 Hz, and 2000 Hz.
7. Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
8. Pain or discomfort in the ear.

CAUTION: The maximum output of the sound generator falls into the range that can cause hearing loss according to OSHA regulations. In accordance with NIOSH recommendations, the user should not use the sound generator for more than eight (8) hours a day when this is set to a level of 85 dB SPL or above. When the sound generator is set to levels of 90 dB SPL or above the user should not use the sound generator for more than two (2) hours per day. In no case should the sound generator be worn at uncomfortable levels.

Tinnitus Breaker Pro precautions

1. Should the user develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function or increase in tinnitus perception, the user should discontinue use of the sound generator and seek medical evaluation.
2. Discontinue use of the sound generator and consult promptly with a licensed physician if you experience one of the following conditions:
   a. Visible congenital or traumatic deformity of the ear.
   b. History of active drainage from the ear within the previous 90 days.
   c. History of sudden or rapidly progressive hearing loss within the previous 90 days.
   d. Acute or chronic dizziness.
   e. Unilateral hearing loss of sudden or recent onset within the previous 90 days.

f. Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.

g. Pain or discomfort in the ear.

3. Discontinue use of the sound generator and consult promptly with your hearing care professional, if you experience changes in the tinnitus perception, discomfort or interrupted speech perception, while using the Tinnitus Breaker Pro.
4. The volume control is a feature in the TBP module used for adjusting the sound generator output level.To prevent unintended usage by paediatric or physically or mentally disabled users, the volume control must be configured to only provide a decrease of the sound generator output level.
5. Children and physically or mentally disabled users will require guardian supervision while wearing the TBP hearing aid.
6. Adjustment of the Tinnitus Breaker Pro settings, using a smartphone app, should only be performed by the parent or legal guardian in cases where the user is minor.Use of the Beltone Remote Care for remote settings of the tinnitus sound generator, should only be performed by the parent or legal guardian in cases where the user is minor.

Regulatory information

Warranties and repairs

The manufacturer provides a warranty on hearing aids in the event of defects in workmanship or materials, as described in applicable warranty documentation. In its service policy, the manufacturer pledges to secure functionality at least equivalent to the original hearing aid. As a signatory to the United Nations Global Compact initiative, the manufacturer is committed to doing this in line with environment-friendly best practices. Hearing aids therefore, at the manufacturer's discretion, may be replaced by new products or products manufactured from new or serviceable used parts, or repaired using new or refurbished replacement parts. The warranty period of hearing aids is designated on your warranty card, which is provided by your hearing care professional.

For hearing aids that require service, please contact your hearing care professional for assistance.

Hearing aids that malfunction must be repaired by a qualified technician. Do not attempt to open the case of hearing aids, as this will invalidate the warranty.

Temperature test, transport and storage information

Our hearing aids are subjected to various tests in temperature and damp heating cycling between -25 °C (-13 °F) and +70 °C (+158 °F) according to internal and industry standards.

During normal operation the temperature should not exceed the limit values of 0 °C (+32 °F) to +45 °C (+113 °F) at a relative humidity of 90%, non-condensing. An atmospheric pressure between 500 hPa and 1100 hPa is appropriate.

NOTE: During use, your hearing aids may reach temperatures up to +43 °C (+109 °F).

During transport or storage, the temperature should not exceed the limit values of -20 °C (-4 °F) to +60 °C (+140 °F) at a relative humidity of 90% RH, non-condensing (for a limited time).

Statement

This device complies with part 15 of the FCC rules and ISED rules. Its operation is subject to the following two conditions:

1. This device may not cause harmful interference.
2. This device must accept any interference received, including interference that may cause undesired operation.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC rules and ISED rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used according to the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

– Reorient or relocate the receiving antenna
– Increase the separation between the equipment and the receiver

– Connect the equipment to an outlet or a circuit that is different from the one to which the receiver is connected
– Consult the dealer or an experienced radio/TV technician for help.

Changes or modifications can void the user's authority to operate the equipment.

The products are in compliance with the following regulatory requirements:

- In the EU: The device conforms to the Essential Requirements according to Annex I of Council Directive 93/42/EEC for medical devices (MDD).
- Hereby, Beltone A/S declares that the radio equipment types BEB60 and BEB70 are in compliance with Directive 2014/53/EU.
- The full text of the EU declaration of conformity is available at the following internet address: www.beltone.com/en/declarations.
- In the US: FCC CFR 47 Part 15, subpart C.
- Other identified applicable international regulatory requirements in countries outside the EU and US. Please refer to local country requirements for these areas.
- In Canada: these hearing aids are certified under the rules of ISED.
- Japanese Radio Law and Japanese Telecommunications Business Law Compliance: This device has been certified pursuant to the Japanese Radio Law (電波法) and the Japanese Telecommunications Business Law (電気通信事業法). This device should not be modified (otherwise the granted designation number will become invalid).

Type designations

Hearing aid type designations for models included in this user guide are:

BEB60, FCC ID: X26BEB60, IC: 6941C-BEB60; **BEB70**, FCC ID: X26BEB70, IC: 6941C-BEB70.

This device includes an RF transmitter which operates in the frequency band of 2.4 GHz – 2.48 GHz.

Hearing aid variants

Some models may not be available in your region.

Mini Behind-the-ear (BTE) hearing aids of type **BEB60** with FCC ID X26BEB60, IC number 6941C-BEB60 and size 312 battery are available in the following variants:

RLY466-DW, RLY366-DW, RLY266-DW

Nominal RF output power transmitted is: -1 dBm.

Behind-the-ear (BTE) hearing aids of type **BEB70** with FCC ID X26BEB70, IC number 6941C-BEB70 and size 13 battery are available in the following variants:

RLY476-DW, RLY376-DW, RLY276-DW, RLY176-DW

Nominal RF output power transmitted is: -2 dBm.

Symbols

	WARNING: Points out a situation that could lead to serious injuries.		Legal manufacturer.
	CAUTION: Indicates a situation that could lead to minor and moderate injuries.		Date of manufacture.
	Follow instructions for use. (Logo in blue)		
	Product is a Type B applied part.	UDI	Unique Device Identification.
	Equipment includes an RF transmitter.		IMDA label for Singapore.
	Regulatory compliance mark for Australia and New Zealand.	℞ ONLY	By prescription only (US).
	Do not dispose of electronic devices or batteries with ordinary household waste. They should be disposed of at sites intended for electronic waste or returned to your hearing care professional for safe disposal. Please note: There may be specific regulations in your country.		

Technical specifications

Mini BTE

Models: RLY466-DW, RLY366-DW, RLY266-DW

		Thin tube	Closed	
Reference test gain (60 dB SPL input)	HFA	36	39	dB
Full-on gain (50 dB SPL input)	Max. HFA	52 47	56 49	dB
Maximum output (90 dB SPL input)	Max. HFA	123 113	121 116	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz 3200 Hz	0.4 0.1 0.4 0.2	0.7 0.6 0.6 0.1	%
Telecoil sensitivity (1 mA/m input) HFA - SPLIV @ 31.6 mA/m (ANSI) Full-on telecoil sensitivity @ 1 mA/m	Max. HFA HFA	81 96 77	85 99 79	dB SPL
Equivalent input noise, w/o noise reduction 1/3 Octave Equivalent input noise, w/o noise reduction		22 10	23 10	dB SPL
Frequency range IEC 60118-0: 2015		100-7680	100-6800	Hz
Current Drain (Quiescent/Operating)		1.17/1.22	1.18/1.34	mA





Black curve: Thin tube;
Grey curve: Closed dome.

Data in accordance with ANSI S3.22-2014, IEC 60118-0:2015. Measured in a 2cc coupler.

BTE

Models: RLY476-DW, RLY376-DW, RLY276-DW, RLY176-DW

		Thin tube	Closed	
Reference test gain (60 dB SPL input)	HFA	40	45	dB
Full-on gain (50 dB SPL input)	Max. HFA	52 49	57 52	dB
Maximum output (90 dB SPL input)	Max. HFA	128 117	126 122	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz 3200 Hz	0.5 0.1 0.6 0.2	0.7 0.9 0.6 0.2	%
Telecoil sensitivity (1 mA/m input) HFA - SPLIV @ 31.6 mA/m (ANSI) Full-on telecoil sensitivity @ 1 mA/m	Max. HFA HFA	83 101 79	88 105 83	dB SPL
Equivalent input noise, w/o noise reduction 1/3 Octave Equivalent input noise, w/o noise reduction		22 10	22 11	dB SPL
Frequency range IEC 60118-0: 2015		100-7130	100-6170	Hz
Current Drain (Quiescent/Operating)		1.18/1.2	1.2/1.29	mA





Black curve: Thin tube;
Grey curve: Closed dome.

Data in accordance with ANSI S3.22-2014, IEC 60118-0:2015. Measured in a 2cc coupler.

Additional information

Acknowledgements

Portions of this software are written by Kenneth MacKay (micro-ecc) and licensed under the following terms and conditions:

Copyright ® 2014, Kenneth MacKay. All rights reserved.

Redistribution and use in source and binary forms, with or without modification, are permitted provided that the following conditions are met:

- Redistributions of source code must retain the above copyright notice, this list of conditions and the following disclaimer.
- Redistributions in binary form must reproduce the above copyright notice, this list of conditions and the following disclaimer in the documentation and/or other materials provided with the distribution.

THIS SOFTWARE IS PROVIDED BY THE COPYRIGHT HOLDERS AND CONTRIBUTORS “AS IS” AND ANY EXPRESS OR IMPLIED WARRANTIES, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE DISCLAIMED. IN NO EVENT SHALL THE COPYRIGHT HOLDER OR CONTRIBUTORS BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO, PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY

WAY OUT OF THE USE OF THIS SOFTWARE, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE.

NOTE: Use of the Made for Apple badge means that an accessory has been designed to connect specifically to iPhone, iPad, and iPod touch models, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards.

© 2025 GN Hearing A/S. All rights reserved. Beltone is a trademark of GN Hearing A/S. Apple, the Apple logo, iPhone, iPad, iPod touch are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc., registered in the US and other countries. Android, Google Play and the Google Play logo are trademarks of Google LLC. The Bluetooth word mark and logos are registered trademarks owned by Bluetooth SIG, Inc.

Notes

Notes

Notes

Manufacturer according to EU Medical Device Directive 93/42/EEC:

Beltone A/S
Lautrupbjerg 7
DK-2750 Ballerup
Denmark

Local contacts:

United Kingdom
GN Hearing UK Ltd
Unit 13 Talisman Business Centre
Bicester OX26 6HR
United Kingdom
Tel.: +44 1869 352 800
beltone.com

Australia
GN Hearing Australia Pty Ltd
Suite 1001, Level 1
1 Julius Avenue
North Ryde, NSW 2113
Australia
Tel.: (free) 1800 658 955
beltone.com

New Zealand
GN Hearing New Zealand Limited
Ground Floor, North Entrance
4 Fred Thomas Drive
Takapuna, Auckland, 0622
Tel.: (free) 0800 900 126
beltone.com

CE 0297

Any serious incident that has occurred in relation to the device should be reported to the Legal manufacturer Beltone A/S and the competent authority of the EU Member State in which the user and/or patient is established.

401381000GB- 25.11 - Rev. I (Date of issuance: 2025.11.01)